

KASP thermal cycling conditions

68-62 °C touchdown protocol

This alternative thermal cycle programme is typically used for assays with high percentage GC content.

Protocol stage	Temperature	Duration	Number of cycles for each stage
Stage 1 Hot-start <i>Taq</i> activation	94 °C	15 minutes	x 1 cycle
Stage 2 Touchdown	94 °C	20 seconds	x 10 cycles
	68 °C (68 °C decreasing 0.6 °C per cycle to achieve final annealing/extension temperature of 62 °C)	60 seconds	
Stage 3 Amplification	94 °C	20 seconds	x 26 cycles
	62 °C	60 seconds	
Optional stage 4 (read stage for qPCR instruments only)	30 °C (any temperature below 40 °C is suitable for the read stage)	60 seconds	x 1 cycle

* Please note that stage 4 of the above program is only required if running and reading KASP™ genotyping reactions on a qPCR machine. If running the KASP thermal cycle program on a Peltier block or a Hydrocycler, only Stages 1, 2 and 3 are needed although you must ensure that the reaction plates are cooled to <40 °C before performing the template read.

Explanation of touchdown PCR

KASP chemistry utilises a two-step touchdown PCR method, with the elongation and annealing steps incorporated into a single step. The temperature used for the annealing stage determines the specificity of the reaction and hence the ability of the primers to anneal to the DNA template. A touchdown PCR involves starting with a high annealing temperature and incrementally decreasing the annealing temperature each PCR cycle. The higher annealing

temperatures in the early cycles of a touchdown ensure that only very specific base pairing will occur between the DNA and the primer, hence the first sequence to be amplified is most likely to be the sequence of interest. The annealing temperature is gradually decreased to increase the efficiency of the reaction. The regions that were originally amplified during the highly specific early touchdown cycles will be further amplified and outcompete any non-specific amplification that may occur at the lower temperatures.

@LGCBiosearch | biosearchtech.com

For any queries about running KASP reactions in your laboratory please contact the technical support team at techsupport@lgcgroup.com.

All trademarks and registered trademarks mentioned herein are the property of their respective owners. All other trademarks and registered trademarks are the property of LGC and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details. No part of this publication may be reproduced or transmitted in any form or by any means, electronic or mechanical, including photocopying, recording or any retrieval system, without the written permission of the copyright holder. © LGC Limited, 2023. All rights reserved. GEN/1087/EK/0323

BIOSEARCH™
TECHNOLOGIES
GENOMIC ANALYSIS BY LGC

For Research Use Only. Not for use in diagnostic procedures.